

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 21 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130174</u>
Name <u>Gary Lowe</u>		ID# <u>7219</u> Date <u>5/19/09</u>
A	Agency <u>Haines Police</u>	Phone # <u>766-2121</u>
Instrument Location <u>Haines PD</u>		
Alco S/N <u>X124631</u> ✓ Target Value <u>.082</u> ✓ High Pressure <u>700</u>		
B	Alco Test Values <u>.082</u> ✓ 1 <sup>st</sup> Alco	
	<u>.082</u> ✓ 2 <sup>nd</sup> Alco	
Signature <u>[Signature]</u>		
(OVER)		

BMB  
5/22/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 27<sup>th</sup> day of May, 2009.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130174

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

BMB  
5/22/09

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

NOV 19, 2009

OPERATOR'S NAME:

LOVE, GARY A

OPERATOR'S NUMBER: 7219

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/NI:

TEST

D.L. #: 12345

DEPT/AGENCY: HNS1

CASE/REPORT: 090519

TEST TYPE: U

ALCO TARGET VALUE: .002

ALCO S/N: X124631

— BREATH ANALYSIS —

.002 ADJUSTED FOR 30.17 in

ALCO TARGET	.002	09:32
BLANK TEST	.000	09:33
INTERNAL STANDARD	VERIFIED	09:33
ALCO TU 30.17 in	.002	09:33
BLANK TEST	.000	09:34
SUBJECT SAMPLE	.000	09:35
BLANK TEST	.000	09:35
ALCO TU 30.17 in	.002	09:36
BLANK TEST	.000	09:36

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130174

NOV 19, 2009

TIME 09:37

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 30.17 in

FLOW DETECTOR: OKAY

PUMP

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z { | } ~ ¡ ¢